

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

FEB 08 2010

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N <u>130125</u>	
Name <u>Anthony M. Wiles</u>		ID# <u>6877</u>	Date <u>1-22-2010</u>
A	Agency <u>AST - D Detachment</u>		Phone # <u>451-5100</u>
Instrument Location <u>1979 Peger Rd., FBKS AK 99701</u>			
B	Alco S/N <u>X172971</u>		Target Value <u>.082</u> High Pressure <u>1200 psi</u>
Alco Test Value Average <u>.079</u>		<u>.079</u>	
1 st Alco		2 nd Alco	
Signature <u>Anthony Wiles</u>			
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 16th day of Feb., 2010.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130125 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125 ✓
JANUARY 22, 2010 ✓ *BMB 2/11/10*
OPERATOR'S NAME: WILLY JOHNSON
TESTER'S NAME: J. J. J.
SUBJECT'S LAST NAME: J. J. J.
SUBJECT'S FIRST NAME: J. J. J.
SUBJECT'S BIRTH DATE: J. J. J.
SUBJECT'S BIRTH TIME: J. J. J.
O.L. #1 N
TESTER'S SIGNATURE: J. J. J.
TEST TIME: J. J. J.
ALCO TARGET VALUE: 0.00 ✓
ALCO RANGE: 0.00-0.20 ✓
--- BREATH ANALYSIS ---
ALCO ADJUSTED FOR 20°C: 0.00 ✓
ALCO TARGET: 0.00 06:121
BLOW TEST: 0.00 06:122
OVERALL STATUS: PASS 06:122
ALCO TEST: 0.00 06:123
BLOW TEST: 0.00 06:124
BLOW TEST: 0.00 06:124
BLOW TEST: 0.00 06:125
BLOW TEST: 0.00 06:125

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125 ✓
JANUARY 22, 2010 ✓
TIME 06:127
--- TEST RESULTS ---
COMPUTER: OK
PROGRAM: OK
SOFTWARE DATE: 01/20/01
SEATING: OK
SAMPLE CHAMBER: 40c
BREATH TUBE: 40c
PAROMETER: 0.00 06:127
FLOW DETECTOR: OK
PUMP: OK
HIGH PRESS: OK
DETECTOR: OK
FILTERS: OK
WASTE AT CAP: OK
CALIBRATION: OK
TESTER'S SIGNATURE: J. J. J.
TESTER'S NAME: J. J. J.
TESTER'S BIRTH DATE: J. J. J.
TESTER'S BIRTH TIME: J. J. J.